

**ZAJEDNIČKI POPIS BRZIH ANTIGENSKIH TESTOVA NA COVID-19
u skladu s preporukama Vijeća EU te sa strategijama testiranja zemalja***

Proizvođač	Komercijalni naziv
AAZ-LMB	COVID-VIRO® Rapid antigen test COVID-19
Abbott Rapid Diagnostics	Panbio™ COVID-19 Ag Rapid Test
Abbott Rapid Diagnostics	Panbio™ COVID-19 Ag Rapid Test
Acon Biotech (Hangzhou) Co., Ltd	FlAwflex SARS-CoV-2 Antigen Rapid Test
AESKU.DIAGNOSTICS GmbH & Co, KG	AESKU.RAPID SARS-CoV-2
AMEDA Labordiagnostik GmbH	AMP Rapid Test SARS-CoV-2 Ag
AMEDA Labordiagnostik GmbH	AMP Rapid Test SARS-CoV-2 Ag
Becton Dickinson	BD Veritor System for Rapid Detection of SARS-CoV-2
Beijing Hotgen Biotech Co., Ltd.	Novel Coronavirus 2019-nCoV Antigen Test (Colloidal Gold)
Beijing Lepu Medical Technology	SARS-CoV-2 Antigen Rapid Test Kit (Colloidal Gold immunochromatography)
Beijing Lepu Medical Technology	SARS-CoV-2 Antigen Rapid Test Kit (Colloidal Gold immunochromatography)
Beijing Wantai Biological Pharmacy Enterprise Co Ltd	WANTAI SARS-CoV-2 Ag Rapid Test (FIA)
BIONOTE	NowCheck® COVID-19 Ag Test
BIOSYNEX SWISS SA	BIOSYNEX COVID-19 Ag BSS
BIOSYNEX SWISS SA	BIOSYNEX COVID-19 Ag BSS
CerTest Biotect S.L.	CerTest SARS-CoV-2 CARD TEST
CerTest Biotect S.L.	CerTest SARS-CoV-2 CARD TEST
Core Technology Co., Ltd	Coretests COVID-19 Ag Test
Dialab	DIAQUICK COVID -19 Ag Cassette
GenBody Inc	GenBody COVID-19 Ag Test
Green Cross Medical Science Corp.	GENEDIA W COVID-19 Ag
Guangdong Wesail Biotech Co. Ltd	COVID-19 AG Test Kit
Guangzhou Wondfo Biotech Co., Ltd	Wondfo 2019-nCoV Antigen Test (Lateral Flow Method)
Hangzhou Clongene Biotech	Clungene COVID-19 Antigen Rapid Test Kit
Hangzhou Clongene Biotech Co., Ltd.	COVID-19 Antigen Rapid Test Kit
Healgen Scientific Limited	Coronavirus Ag Rapid Test Cassette (Swab)
Healgen Scientific Limited	Coronavirus Ag Rapid Test Cassette (Swab)
Humasis Co. Ltd	HUMASIS COVID-19 Ag test
Joinstar Biomedical Technology	COVID-19 Antigen Rapid Test (Colloidal Gold)

Joinstar Biomedical Technology	COVID-19 Antigen Rapid Test (Colloidal Gold)
LumiQuick Diagnostics Inc.	QuickProfile™ COVID-19 ANTIGEN Test
LumiraDX UK LTD	LumiraDx SARS-CoV-2 Ag Test
LumiraDX UK LTD	LumiraDx SARS-CoV-2 Ag Test
MEDsan GmbH	MEDsan® SARS-CoV-2 Antigen Rapid Test
MEDsan GmbH	MEDsan® SARS-CoV-2 Antigen Rapid Test
MP Biomedicals Germany	Rapid SARS-CoV-2 Antigen Test Card
MP Biomedicals Germany	Rapid SARS-CoV-2 Antigen Test Card
Nal von minden GmbH	NADAL COVID -19 Ag Test
nal von minden GmbH	NADAL COVID -19 Ag Test
PCL Inc	PCL COVID19 Ag Rapid FIA
Precision Biosensor Inc (Axon Lab SG)	Exdia COVI-19 Ag Test
Qingdao Hightop Biotech Co Ltd	SARS-CoV-2 Antigen Rapid Test
Quidel Corporation	Sofia 2 SARS Antigen FIA
Quidel Corporation	Sofia 2 SARS Antigen FIA
Rapigen Inc.	BIOCREDIT COVID-19 Ag - SARS-CoV 2 Antigen test
Roche (SD BIOSENSOR)	SARS-CoV-2 Antigen Rapid Test
Safecare Biotech Hangzhou Co	COVID-19 Ag Rapid Test Kit (Swab)
Safecare Biotech Hangzhou Co	COVID-19 Antigen Rapid Test Kit (Swab)
SD BIOSENSOR, Inc.; Roche	STANDARD F COVID-19 Ag FIA
SD BIOSENSOR, Inc.; Roche	STANDARD Q COVID-19 Ag Test
SD BIOSENSOR, Inc.; Roche	STANDARD F COVID-19 Ag FIA
SD BIOSENSOR, Inc.; Roche	STANDARD Q COVID-19 Ag Test
Shenzen Ultra-Diagnostics Biotec Co.	SARS-COV-2 Antigen test Kit (colloidal gold)
Shenzhen Watmind Medical Co., Ltd	SARS-CoV-2 Ag Diagnostic Test Kit (Colloidal Gold)
Siemens Healthineers	CLINITEST Rapid COVID-19 Antigen Test
Siemens Healthineers	CLINITEST Rapid COVID-19 Antigen Test
TODA Pharma	TODA CORONADIAG Ag®
Xiamen Boson Biotech Co	Rapid SARS-CoV-2 Antigen Test card
Xiamen Boson Biotech Co	Rapid SARS-CoV-2 Antigen Test card
Zhejiang Orient Gene Biotech Co., Ltd	Coronavirus Ag Rapid Test Cassette (Swab)
Zhejiang Orient Gene Biotech Co.,Ltd	Coronavirus Ag Rapid Test Cassette (Swab)

* Testovi sa navedene liste udovoljavaju uvjetima kako slijedi: 1. Nose oznaku CE; 2. Ispunjavaju minimalne zahtjeve učinkovitosti od ≥ 90 % osjetljivosti i ≥ 97 % specifičnosti; 3. Potvrdila ih je najmanje jedna država članica kao prikladne za upotrebu u kontekstu bolesti COVID-19, pri čemu je navela pojedinosti o metodologiji i rezultatima takvih studija, tj. koja je vrsta uzorka upotrijebljena za validaciju, pod kojim je uvjetima upotreba testa procijenjena, te je li bilo poteškoća u vezi s traženim kriterijima osjetljivosti ili drugim elementima učinkovitosti.